Automatic High Voltage Breakdown Tester

Model : QHBT-SERIES

Application:

- Insulation, Cables, Transformers Industries
- Electricity Distribution substations
- Component Manufacturers
 Panel Manufacturers
 - urers Electrical test / Calibration Labs



"QUADRANT" make 'Automatic AC/DC HIGH VOLTAGE TEST SET' are ideal to test dielectric strength of

different insulating materials , motor winding transformers, Ceramic HV insulators, cables and different electrical

components. It is useful in transformer manufacturing industries, HV motor manufacturing units, cables and

insulator industries etc.

TECHNICAL SPECIFICATIONS

INPUT : 415 V, 2ph.(of 3 ph.),50 Hz A.C.OR 230V Single Phase OUTPUT : upto 0 – 100 KV AC/DC Variable LEAKAGE CURRENT : upto 100 mAmp (settable) by rotary switch in steps of 25-50-100 mA DUTY : Intermittent (I.e; 1 Minute Withstand Test)

COOLING : Oil /Air Cooled

ACCESSORIES

- 1. 5 Meter Iput Mains Power Cord
- 2. 5 Meter Interconnecting cable for Testing Panel & High Voltage Transformer
- 3. Rectifier unit for DC output
- 4. 5 Meter Interconnecting cable for Testing Panel& Variable Auto Transformer
- 5. Operating Manual
- 6. Works Test Report
- 7. Warranty Certificate

To keep pace with the latest technology Quadrant Measurements Pvt. Ltd. reserves the right to update it's system. Hence, information in the brochure is likely to change without notice.



QUADRANT MEASUREMENTS Pvt. Ltd. Simplified Measurements Solutions Works : D - 51, MIDC, Satpur, Nashik, Maharashtra, INDIA.

Email : sales@quadrantmeasurements.com Website : www.quadrantmeasurements.com

High Voltage Breakdown Tester

Model : QHBT-SERIES

PROTECTION

- (If use is provided in incoming side.
- (ii) Automatic Tripping Mechanism for protecting
- the H.V. transformer against overloads as well as short circuits. This tripping
- device senses the leakage current to be higher that the set value and trips the H.V. circuit.
- (iii) One Zero-Start Interlock is also provided to
- ensure that the H.V. circuit cannot be energized unless the motorized variable
- auto-transformer is at zero position or brought back to zero position.
- (iv) One Trip Current adjustment switch (in steps)
- is also provided to adjust the tripping current.

SETUP CONFIGURATION

It generally consists of two units namely

- (i) Control Unit
- (ii) HV Unit
- The powder coated sturdy steel

cabinet of the 'Control Unit' houses meters, switches, indicating lamps, push buttons, fuses etc. which are fitted on the top panel of the cabinet. Along with these, Contactor, protection CT, wirings, one continuously variable auto transformer to vary the output voltage of suitable capacity are also mounted inside the cabinet. One single/two phase (w.r.t. L.T.) copper double wound H.V. step-up transformer comprises of the H.V. unit. The transformer will be oil natural cooled and housed in an insulated cylindrical jacket which is placed on M.S. Trolley. A metallic sheet is placed on the top of the insulated cylindrical jacket through which an H.V. terminal is brought out. L.T terminals and starting end of H.V. coil is brought out at the bottom side of the M.S. Trolley marked as " L1, L2 & V "respectively just below the insulated cylindrical jacket. The control unit is separately housed in M.S. cabinet. All nomenclatures are engraved in anodized Aluminum plates which are placed on the cabinet. 4 nos. casting wheels are provided at the bottom of the cabinet for easy mobility.

To keep pace with the latest technology Quadrant Measurements Pvt. Ltd. reserves the right to update it's system. Hence, information in the brochure is likely to change without notice.



