

Technique Data

| • | Measurement mode | Scan mode |
|---|--|--|
| • | Measurement range | 0-100mg/100mlBAC |
| • | Measurement time | Adjustable, Minimum 1 second |
| • | Continuous measuremen | t times One charging, 2000 times |
| • | Display | Figure and graph |
| • | Alarm Red, Green, yellow LED indication, alarm limit can be adjusted | |
| • | Battery | Rechargable Li-ion |
| • | Sensor | Fuel cell |
| • | Mouthpiece | No need |
| • | Power save mode | Automatic switch off after 2.5 minutes without use |
| • | Calibration | 12 months, according to some requirements |
| | | 8 A 76 |

Mahathi Systems India Pvt.Ltd, LGF-29, Sahara Plaza, Sahara States, Mansoorabad, Hyderabad-500068, Cell: 93 92 509 500