

3. Value -Indicating precision :

Standard hardness test block	Max error of the indicated value%	Repetition error of the indicated value%
≤125	±3	3
125 < HBW ≤ 225	±2.5	2.5
> 225	±2	2

The Packing List

Model MHB-3000 Digital Brinell Hardness Tester one

Case for the Accessories of the Main Machine

φ2.5、φ5、φ10mm steel ball indenter		1pc(each)
Big ,small plane testing table, "V" testing table		1pc(each)
Standard hardness test block		
HBW10/3000	(150 ~ 250)	1pc
HBW5/750	(150 ~ 250)	1pc
Spare Fuse(2A)		2pcs
Power cord		1pc
Usage instruction manual		1copy
20× reading microscope		1copy